

PO071-TUE

Reliability of a new point-of-care portable coagulometer for PT-INR test performed in the hospital anticoagulation clinic

Paniccia R1, Priora R2, Marcucci R3, Mannini L2, Poli D2, Tafuro EL3, Attinasi F2, Gori L2, Liotta AA2 and Abbate R3

1Experimental and Clinical Medicine, University of Florence; 2Careggi Hospital; 3University of Florence, Florence, Italy

Background: Recently, a new coagulation point-of-care testing (POCT) system, the microINR portable coagulometer (Instrumentation laboratory-IL, Italy), has been introduced for assessing PT-INR in capillary whole blood samples of patients on oral vitamin K antagonist (VKA) therapy.

Aims: This study was aimed to evaluate the precision and accuracy of this POCT coagulometer.

Methods: Capillary whole blood PT-INR by using POCT device were assessed in 288 patients on oral VKA therapy. At the same time, citrated blood was withdrawn for comparing citrated plasma PT-INR results obtained from a laboratory conventional method (ACL TOP700, IL, Italy).

Results: Significant correlations were observed between POCT and laboratory PT-INR ($\rho = 0.96$, $P < 0.001$, $n = 288$). For all determinations, the agreement analysis by using Bland Altman plot revealed that the mean differences between POCT system and laboratory method was: 0.05 ± 0.4 INR ($P < 0.01$). The 95° limits of agreement ranged from -0.7 to 0.85 INR ($P < 0.001$). Across different POCT PT-INR ranges, the following mean differences were observed: < 2.0 INR ($n = 20$), 0.03 ± 0.24 INR; 2 to 3 INR ($n = 83$), 0.10 ± 0.29 INR; 3 to 4 INR ($n = 83$), 0.09 ± 0.33 INR; > 4 INR ($n = 102$), -0.04 ± 0.44 INR. In addition, for these different ranges significant correlations between POCT and laboratory PT-INR were observed (at least, $P < 0.01$). Regarding between-cartridge imprecision analysis, the coefficient of variation (CV) within day was: $CV=3.2\%$. The Integrated Quality Control for the guaranteed reliability of system did not failed.

Conclusion: These data show that the POCT system microINR provides precise and accurate results that are significantly comparable with plasma laboratory PT-INR.

Disclosure of Interest: None declared.